



Tech Info Library

Apple Basic Color Monitor: Medical Specifications (12/93)

Revised: 12/14/93
Security: Everyone

Apple Basic Color Monitor: Medical Specifications (12/93)

=====

Article Created: 14 December 1993

TOPIC -----

What is the implosion or protection rating on the Apple Basic Color Monitor in regards to complying with the UL 544 (Underwriters Laboratory Specifications)? This knowledge is for the use of the monitor in a medical environment.

DISCUSSION -----

The UL 544 specification is designed specifically for medical equipment that is directly interfaced to a patient (heart monitor, and so on). The Apple Basic Color display does not fall into this category. However, it does meet or exceed the UL478 and UL1950 standards.

Copyright 1993, Apple Computer, Inc.

Tech Info Library Article Number:14228